

इंटरनेट

मानक

Disclosure to Promote the Right To Information

Whereas the Parliament of India has set out to provide a practical regime of right to information for citizens to secure access to information under the control of public authorities, in order to promote transparency and accountability in the working of every public authority, and whereas the attached publication of the Bureau of Indian Standards is of particular interest to the public, particularly disadvantaged communities and those engaged in the pursuit of education and knowledge, the attached public safety standard is made available to promote the timely dissemination of this information in an accurate manner to the public.

“जानने का अधिकार, जीने का अधिकार”

Mazdoor Kisan Shakti Sangathan


“The Right to Information, The Right to Live”

“पुराने को छोड़ नये के तरफ”

Jawaharlal Nehru

“Step Out From the Old to the New”


IS 7244 (1974): Thermometer for Mercury Barometer [PGD 21: Meteorological Instruments]



“ज्ञान से एक नये भारत का निर्माण”

Satyanarayan Gangaram Pitroda

“Invent a New India Using Knowledge”



“ज्ञान एक ऐसा खजाना है जो कभी चुराया नहीं जा सकता है”

Bhartrhari—Nitiśatakam

“Knowledge is such a treasure which cannot be stolen”

BLANK PAGE



IS : 7244 - 1974

Indian Standard
SPECIFICATION FOR
THERMOMETER FOR MERCURY BAROMETER

UDC 551.508.41 : 536.51



© Copyright 1974

INDIAN STANDARDS INSTITUTION
MANAK BHAVAN, 9 BAHADUR SHAH ZAFAR MARG
NEW DELHI 110001

Price Rs 2.50
Gr 2

September 1974

Indian Standard

SPECIFICATION FOR THERMOMETER FOR MERCURY BAROMETER

Meteorological Instruments Sectional Committee, EDC 69

Chairman

MISS A. MANT

Representing

India Meteorological Department, New Delhi

Members

SHRI N. R. CHAKRABORTY	National Instruments and Ophthalmic Glass Ltd, Calcutta
SHRI P. K. GHOSH (<i>Alternate</i>)	
DIRECTOR (METEOROLOGY)	Indian Navy
STAFF OFFICER (METEOROLOGY) (<i>Alternate</i>)	
JOINT DIRECTOR, RESEARCH (B & F)	Research, Designs & Standards Organization (Ministry of Railways), Lucknow
DEPUTY DIRECTOR, RESEARCH (B & F) (<i>Alternate</i>)	
SHRI R. H. MENDONSA	Lawrence & Mayo (India) Private Ltd, Bombay
SHRI S. VARADARAJAN (<i>Alternate</i>)	
GP CAPT P. A. MENON	Indian Air Force
WG CDR R. S. CHHATWAL (<i>Alternate</i>)	
SHRI D. D. PURI	Central Scientific Instruments Organization (CSIR), Chandigarh
SHRI A. N. AGARWAL (<i>Alternate</i>)	
SHRI S. M. RAZVI,	Director General, ISI (<i>Ex-officio Member</i>)
Deputy Director (Mech Engg)	
(<i>Secretary</i>)	

© Copyright 1974

INDIAN STANDARDS INSTITUTION

This publication is protected under the *Indian Copyright Act* (XIV of 1957) and reproduction in whole or in part by any means except with written permission of the publisher shall be deemed to be an infringement of copyright under the said Act.

Indian Standard

SPECIFICATION FOR THERMOMETER FOR MERCURY BAROMETER

0. FOREWORD

0.1 This Indian Standard was adopted by the Indian Standards Institution on 31 January 1974, after the draft finalized by the Meteorological Instruments Sectional Committee had been approved by the Mechanical Engineering Division Council.

0.2 The mercury barometer is commonly used in India for the measurement of atmospheric pressure for meteorological purposes. Since the barometer reading is a function of its temperature, a thermometer is an essential adjunct to a mercury barometer.

0.3 An Indian Standard for mercury barometers has already been published (*see* IS : 5798-1970*). The formulation of a similar standard for thermometers for use with mercury barometers has, therefore, become necessary.

0.4 In the formulation of this standard, due consideration has been given to the requirements laid down by the World Meteorological Organization, Geneva, in addition to the special requirements obtaining in this country.

0.5 For the purpose of deciding whether a particular requirement of this standard is complied with, the final value, observed or calculated, expressing the result of a test, shall be rounded off in accordance with IS : 2-1960†. The number of significant places retained in the rounded off value should be the same as that of the specified value in this standard.

1. SCOPE

1.1 This standard covers the requirements for thermometers for use with the mercury barometers conforming to IS : 5798-1970*.

*Specification for mercury barometers.

†Rules for rounding off numerical values (*revised*).

2. TERMINOLOGY

2.1 For the purpose of this standard the definitions given in IS : 2627-1963* shall apply.

3. TYPE

3.1 The thermometer shall be of the liquid-in-glass type, having a solid stem and graduated for vertical total immersion. It shall have a range from -20 to $+50^{\circ}\text{C}$.

4. MATERIAL

4.1 Glass Tubing — The glass capillary and bulb tubing used for the thermometer shall be as prescribed in IS : 4610-1968†.

4.2 Thermometric Liquid — Recommended thermometric liquid is pure and dry mercury, preferably alloyed with 8.5 percent of thallium by mass.

5. DIMENSIONS

5.1 The thermometer shall conform to the dimensions shown in Fig. 1.

6. GENERAL REQUIREMENTS

6.1 Construction

6.1.1 The thermometer shall conform to the shape as shown in Fig. 1. The stem shall be straight, but for a bend near the bulb displacing the centre of the bulb by 4 mm towards the barometer.

6.1.2 No enlargement of the bore shall be permissible in the graduated portion of the stem or within 10 mm of either end of the scale.

6.2 Bulb

6.2.1 The bulb shall be cylindrical. The shape and finish of the bulb shall be such as not to entrap the thermometric liquid.


6.2.2 Top Finish — The top of the thermometer shall have a bent nib as shown in Fig. 1.

6.2.3 Expansion Chamber — The thermometer shall be so constructed as to withstand a temperature of 65°C without damage. An elongated or pear shaped expansion chamber with a hemispherical top and without re-entrant shoulders shall be provided at the top end of the capillary in line with the capillary bore to enable the thermometer to withstand the specified range of temperature.

6.2.4 The thermometer shall be suitably annealed before engraving.

*Glossary of terms relating to liquid-in-glass thermometers.

†Specification for glass tubes for general purpose and reference thermometers.



Length of Graduation

1°C	...	3 mm
5°C	...	5 mm
10°C	...	5 mm

All dimensions in millimetres.

FIG. 1 DIMENSIONS FOR THERMOMETER FOR MERCURY BAROMETER

7. GRADUATION AND FIGURING

7.1 The graduation lines shall be clearly engraved on the stem at every full degree Celsius and shall be of uniform thickness not exceeding 0.15 mm. They shall be filled with black pigment.

7.2 The graduation lines shall be at right angles to the axis of the thermometer when viewed from the front, keeping the thermometer in a vertical position. The graduations shall all finish on an imaginary line parallel to the axis on the right-hand side, when viewed from the front as shown in Fig. 1.

7.3 The numerals -20, -10, 0, 10, 20, 30, 40 and 50 shall be etched at the right places on the stem on the right-hand side as shown in Fig. 1.

7.4 The figures shall be placed in such a way that they are bisected by an extension of the line to which they refer; alternately, they shall be placed immediately above the extended line.

7.5 The number of graduations in a length of 20 mm shall not exceed 10.

8. ACCURACY

8.1 Scale Error — The maximum permissible scale error at any point shall be within $\pm 0.2^{\circ}\text{C}$.

8.2 Interval Error — The change of error between two points separated by an interval of 10°C on the stem shall not exceed 0.1°C .

9. MARKING AND PACKING

9.1 Marking

9.1.1 Each thermometer shall be marked legibly with the following information:

- a) The letter 'C' near the top of the scale;
- b) Maker's name or recognized trade-mark, if any, at the back of the thermometer; and
- c) Serial number of the lot and year of manufacture.

9.1.2 The thermometer may also be marked with the ISI Certification Mark.

NOTE — The use of the ISI Certification Mark is governed by the provisions of the Indian Standards Institution (Certification Marks) Act and the Rules and Regulations made thereunder. The ISI Mark on products covered by an Indian Standard conveys the assurance that they have been produced to comply with the requirements of that standard under a well-defined system of inspection, testing and quality control which is devised and supervised by ISI and operated by the producer. ISI marked products are also continuously checked by ISI for conformity to that standard as a further safeguard. Details of conditions under which a licence for the use of the ISI Certification Mark may be granted to manufacturers or processors, may be obtained from the Indian Standards Institution.

9.2 Packing — Each thermometer shall be wrapped in a piece of thin tissue paper and packed in a circular cardboard carton; alternately, they shall be suitably packed as agreed to between the purchaser and the supplier.

10. TESTING AND INSPECTION

10.1 Each thermometer shall comply with all the requirements of this standard. The calibration test shall be carried out in accordance with IS : 6274-1971*.

*Method of calibrating liquid-in-glass thermometers.

INDIAN STANDARDS

ON

METEOROLOGICAL INSTRUMENTS

IS:

4849-1968	Rain measures
5225-1969	Raingauge, non-recording
5235-1969	Raingauge, recording
5793-1970	Aneroid barometers
5798-1970	Mercury barometers
5799-1970	Windvane
5900-1970	Hair hygrograph
5901-1970	Thermograph, bimetallic
5912-1970	Anemometer, cup counter
5924-1970	Clock mechanisms and drum for meteorological instruments
5945-1970	Barograph, aneroid
5946-1970	Whirling psychrometers
5947-1970	Charts for recording meteorological instruments
5948-1970	Thermometer screens
5973-1970	Pan evaporimeter
6805-1973	Assmann psychrometer
6806-1973	Snowgauge
6871-1973	Distant indicating wind equipment

INDIAN STANDARDS INSTITUTION

Manak Bhavan, 9 Bahadur Shah Zafar Marg, NEW DELHI 110001

Telephone : 27 01 31 (20 lines)

Telegrams : Manaksanathe

Branch Offices:

	Telephone
'Sadhna', Nurmohamed Shaikh Marg, Khanpur, AHMEDABAD 380001	2 03 91
F Block, Unity Bldg, Narasimharaja Square, BANGALORE 560002	2 76 49
534 Sardar Vallabhbhai Patel Road, BOMBAY 400007	35 69 44
5 Chowringhee Approach, CALCUTTA 700013	23-08 02
Kothi No. 90, Sector 18 A, CHANDIGARH	—
5-8-56/57 Nampally Station Road, HYDERABAD 500001	4 57 11
117/418 B Sarvodaya Nagar, KANPUR 208005	82 72
54 General Patters Road, MADRAS 600002	8 37 81
B. C. I. Bldg (Third Floor), Gandhi Maidan East, PATNA 800004	2 56 55